



**AZ Instrument Corp.**

No.3-2, Chien-Kuo Road, TEPZ Tantz, Taichung 427, Taiwan

Web site: <http://www.az-instrument.com.tw>

E-mail: [info@az-instrument.com.tw](mailto:info@az-instrument.com.tw)

衡欣實業股份有限公司

台中縣潭子鄉建國路 3-2 號

Tel:+886-4-25326668

Fax:+886-4-25326593

---

## **DECLARATION OF IDENTICAL TO TYPE**

**Hereby we**

**AZ INSTRUMENT CORPORATION**

No. 3-2, Chien Kuo Road, TEPZ Tantz, Taichung, Taiwan R.O.C.

Tel:886-4-25326668

Fax:886-4-25326593

**declare that the equipment**

Brand name : **AZ**

New model number of product : **86023**

**is structurally identical to the basic model**

**Model No. : AZ brand 86022**

**TEST REPORT NUMBER: W6M22304-22597-E-11**

(reference no. of test report)

**Authorized Signature**



-----



Worldwide Testing Services (Taiwan) Co., Ltd.

ACCREDITED TEST HOUSE

## CERTIFICATION OF CONFORMITY

Under EU EMC - DIRECTIVE 2014/30/EU –  
Under UK EMC - Electromagnetic Compatibility Regulations 2016

This certifies that the following designated product

### OPTICAL DISSOLVED OXYGEN METER

Model No. : 86022

Brand Name : AZ

.....  
(Product identification)

Has been tested in accordance to essential protection requirements of Council Directive 2014/30/EU, Electromagnetic Compatibility Regulations 2016 and found the test results indeed meet the limitation of the relevant test standard(s) listed below:

**EN 61326-1 (2013)/BS EN 61326-1 (2013),  
(IEC 61000-4-2 Ed.2.0 (2008)/-3 Ed.4.0 (2020)/-8 Ed.2.0 (2009))**

.....  
(Identification of regulations / standards)

This certificate is issued for

**AZ Instrument Corp.**

**No.3-2, Jianguo Rd., Tanzi Dist., Taichung City 427, Taiwan R.O.C.**

.....  
(Name / Address)

#### SPECIAL STATEMENT:

THE CERTIFICATION IS VALID ONLY IN CONNECTION TO THE TEST REPORT NO. W6M22304-22597-E-11 AND TO THE SAMPLE HAS BEEN TESTED BY WORLDWIDE TESTING SERVICES (TAIWAN) CO., LTD.



May 30, 2023

.....  
(Date)

Rex Kao

.....  
Rex Kao, Laboratory Director